



Intended use:

This product is designed for use under recommendations from your health care professional or sleep therapist. This product is used on adult patients (weight above 30kg) whom CPAP or bi-level therapy has been prescribed. The product is re-usable for single patient in medical institutions, home care and sleep treatment center.





Instructions for use:

- Inspect the packaging and product appearance before using. Please do not use if there are any worn or damaged parts.
- Take out the mask set and L-type connector from the box and snap the L-type connector into the mask.
- > See headgear assembly diagram, hook headgear point A to the mask frame connecting rings on both sides, and then bond the two B points.
- ➤ Hold the mask over nose and mouth, and then position the headgear.
- > Fast release buckle to the mask frame Quick release buckle connection column.
- Adjust the head straps and the hook and loop of the mask comfortably.
- ➤ L-type connector side ISO 4.3mmOD hole for measurement monitoring use.
- ➤ After wearing, connect the ventilation tube connector to the ISO 22mm OD for use.
- Quick release buckle on both sides can be easily removed, also be connected easily.

♦ Cleaning :

- This product is reusable for single patient. \diamond
- Follow the procedures for cleaning/sterilization in medical institutions. \diamond
- Wash the headgear with warm water and neutral detergent or soap. After rinse, air-dry the strap and keep away from direct sunlight or the dryer.
- Mask and L-type connector can be disconnected for cleaning. \diamond
- \diamond Wash with warm water and neutral detergent or soap.
- \diamondsuit After rinse, dry in a cool place. Do not put it in the dryer.
- The uncovered can be autoclaved 25 times with the 15minute*121°C high temperature steam (pressure) sterilization method.



Warning:

- Multiple patients reuse will result in cross-infection of disease.
- Prohibit the use of if the patient has the following symptoms: vomiting, facial trauma burns, impaired myocardial function, Gastroesophageal Reflux Disease, abnormal cough reflex, unconsciousness.
- To be worn by professional medical staff guidance.
- Do not cover or block anti-asphyxiation holes and breathing valve.
- Δ Do not connect worn breathing apparatus has not been started before operation.
- Δ If you feel unwell, STOP using it immediately and seek medical advice immediately.
- Do not use alcohol bleach or solvents containing chlorine and other components to clean the mask set.
- \triangle Do not put the mask set and headgear in the place of the sun exposure or high temperature
- \triangle Do not dismantling the silicone mask and, this action will mask a small loss of functionality, and reduce product life.
- Occur in a leak, slight tightening of the headgear, no improvement in the condition of leakage and then gently tighten the headgear, do not force too much cause discomfort.
- \triangle Dispose of it according to the local medical waste disposal regulations.

Specifications and materials:

- Double Layer Mask size: S/M/L.
- L-type connector: 22mmOD meet ISO 5356-1.
- Headgear : General.


Mask size	LARGE				MEDIUM			SMALL	
Dead space	172 mL				164 mL			152 mL	
Therapy pressure	4 to 20 cm H ₂ O								
Sound power	Average: 32.3 db(A)								
Sound pressure	Average: 25.5 db(A)								
Connector	22mmOD meet ISO 5356-1								
Measuring hole	4.3mmOD meet EN 20594-1								
Resistance	50 L/min: 0.22 cmH ₂ O, 100 L/min: 0.46 cmH ₂ O								
Exhaust Flow	Pressure (cmH ₂ O)			5	10	15	20	25	30
Rate at Given	Flow (L/min)			23	35	47	56	68	80
Therapy	Pressure-flow curve								
Pressure		80							
	Flow(L/min)	60							
		40							
		20							
		0 -							
		4	l	5	10 Pressure	15 (cmH2O)	20	25	30











Fax: +(32)2.732.60.03 E-Mail: mail@obelis.net

